

# FLOW METER REGULATION SYSTEM - PATENT PENDING

## MEDPLUS2



Technical Data	MEDPLUS2
CAPACITA MASSIMA DELL'AMPOLLA MAXIMUM PHIAL CAPACITY CAPACITE MAXIMALE DE L'AMPOULE AMPULLENHCHSTKÖPAZITÄT CAPACIDAD MAXIMA DE LA AMPOLLA	7 ml
PESO DELL'APPARECCHIO WEIGHT OF THE APPARATUS POIDS DE L'APPAREIL GEWICHT DES GERÄTS PESO DEL EQUIPO	7200 g
DIMENSIONI DELL'APPARECCHIO DIMENSIONS OF THE APPARATUS DIMENSIONS DE L'APPAREIL ABMESSUNGEN DES GERÄTS DIMENSIONES DEL EQUIPO	(L) 280 mm (H) 250 mm (P) 280 mm
PRESSIONE MAX MAXIMUM PRESSURE PRESSION MAXIMUM HÖCHSTDRUCK PRESIÓN MÁXIMA	2.4 bar
FLUSSO MAX MAXIMUM FLOW FLUX MAXIMUM HÖCHSTFLUSS FLUJO MÁXIMO	17 l/min.
RUMOROSITA' MAX. NOISE LEVEL BRUIT MAX MAX.GERÄUSCHPEGEL NIVEL DE INTENSIDAD ACÚSTICA	54 dBa approx. (50 cm)
CONDIZIONI D'IMPIEGO CONDITIONS FOR USE CONDITIONS D'EMPLOI ANWENDUNGSBEDINGUNGEN CONDICIONES DE USO	240 min. ON 30 min. OFF

● COMPLIES WITH EUROPEAN STANDARD EN 13544-1

With 10 kit Andy flow nebulizer  
 Spares 4 filters included

## Professional Model

3 YEARS WARRANTY

● ● ● ●  
 Included Accessories



● ● ● ●  
 Nebulizer Performance

Pispor	Valve	OUTPUT RATE [ml/min.]	MMAD [µm]	GSD
A	Open	0.25	5.1	2.6
B	Open	0.28	4.2	2.5
	Closed	0.19	3.6	2.5
C	Open	0.26	3.4	2.5
	Closed	0.16	2.0	2.4

Output tested with a 2.5 ml NaF solution 1%  
 (in compliance with annex CC-EN 13544-1)  
 Ventures & Consultancy Bradford 27th Jan.2004 Bradford University England)

Plant:  
 Via dell'Artigianato 23/25 25080 PADENGHE SUL GARDA (BS) ITALY



Commercial Office:  
 Piazza G. Pasolini 2 - 20159 MILANO ITALY  
 Phone: + 39 02 69301190  
 E mail: medexport@med2000.it

Fax: + 39 02 69006515  
 www.med2000.it